

Technical report (Original)

Optimization of a low-dose 320-slice multi-detector computed tomography chest protocol using a phantom

Chatnapa Nuntue, Anchali Krisanachinda, Kitiwat Khamwan

Department of Radiology, Faculty of Medicine, Chulalongkorn University and King Chulalongkorn Memorial Hospital, Thai Red Cross Society, Bangkok 10330, Thailand

Background: Radiation exposure because of medical imaging is a public safety concern.

Objective: To optimize a protocol for multi-detector computed tomography (MDCT) of the chest with pulmonary nodules based on targeted standard deviation (SD) to achieve minimal patient radiation exposure, while maintaining acceptable diagnostic information.

Methods: A Lungman chest phantom with 5 spheres simulating nodules of 12, 10, 8, 5, and 3 mm diameters at 100 Hounsfield Units (HU) was scanned by 320-slice MDCT by varying targeted SD of 9, 14, 20, and 25, beam pitch of 0.637, 0.813, 1.388, at 120 and 100 kVp, and 10-400 mA. We measured the radiation doses in terms of corresponding volume computed tomography dose index ($CTDI_{vol}$). Quantitative image quality was determined by the percent contrast-to-noise ratio (%CNR). Two independent radiologists evaluated images for nodule detection acceptability on a 5-point rating scale.

Results: We found targeted SD increased results in decreased radiation dose and %CNR. We found $CTDI_{vol}$ reduction between -40% and -88% compared with the default setting for various targeted SD, beam pitch, and kVp. The %CNR was highest at targeted SD 9. There was good agreement between the image quality scores by 2 independent readers ($k = 0.66$).

Conclusion: Targeted SD 20, 10-400 mA at pitch 0.813 and 120 kVp on a nodule size of 5 mm was considered the optimal protocol for low-dose chest CT. Our recommended protocol can reduce the radiation dose substantially under the manufacturer's default protocol, while preserving acceptable image quality for lung nodule detection.

Keywords: CT chest, CT noise index, MDCT, radiation dose reduction, targeted SD

Multi-detector computed tomography (MDCT) is a powerful modality for clinical detection of chest disease especially lung cancers at a smaller size of nodule and earlier stage compared with chest radiography [1, 2]. However, patients undergoing lung screening and patients with the small solitary pulmonary nodules are usually required to follow-up the nodule size several times to evaluate malignancy. This issue has raised concerns for patients owing to the associated radiation exposure and its potential risk of cancer induction [3-5]. At present, the automatic exposure control (AEC) technique has been extensively implemented for reducing the radiation dose in MDCT. Targeted SD is also a part of the AEC system designed by imaging system manufacturers to select the desired image quality by

fixing the image noise level according to the clinical purposes. Low targeted SD indicates higher image quality with excess radiation to patients. Therefore, proper selection of a targeted SD is a crucial part for optimizing the image quality and radiation dose in MDCT. Although previous studies suggest strategies that are used to optimize the radiation dose for MDCT [1-4, 6-12], most studies did not focus on the targeted SD optimization in CT of the chest with lung nodule detection.

This study aimed to optimize the balance between the radiation dose and image quality when varying targeted SD, kVp, and helical pitch in MDCT examinations for pulmonary nodule detection using a lung phantom.

Materials and methods

Chest phantom

This study was approved by the institutional review board (IRB) of the Faculty of Medicine, Chulalongkorn University (certificate of approval No. 140/2013; IRB

Correspondence to: Kitiwat Khamwan, Department of Radiology, Faculty of Medicine, Chulalongkorn University and King Chulalongkorn Memorial Hospital, Thai Red Cross Society, Bangkok 10330, Thailand. E-mail: kitiwat.k@chula.ac.th

No. 078/56). A multipurpose anthropomorphic chest phantom model N1 Lungman (Kyoto Kagaku, Japan) was used to simulate a standard human chest. This phantom is an accurate life-size anatomical model of a human torso. The inner components of the phantom consist of mediastinum, pulmonary vasculature, abdominal block, and synthetic bones that have x-ray attenuation rates relatively to those of human tissues (**Figure 1A**). To mimic pulmonary lesions, five sizes of spherical simulated nodules of 12, 10, 8, 5, and 3 mm in diameters with a CT number of 100 Hounsfield Units (HU) (**Figure 1B**) were attached into the lung field of the chest phantom, i.e. right upper lobe (12 mm), central of left upper lobe (10 mm), right lower lobe (8 mm), 1/3 in peripheral of left lower lobe (5 mm), and central of right middle lobe (3 mm) respectively.

MDCT acquisition protocol and dosimetry

All experiments were conducted with a 320-slice MDCT system (Aquilion ONE, Toshiba Medical Systems) at the Department of Radiology, King Chulalongkorn Memorial Hospital, Thai Red Cross Society, Bangkok, Thailand. The chest phantom was placed and entered into the CT gantry in a head-first supine position, and was acquired from the lung apices to the diaphragm by varying beam pitch of 0.637, 0.813, and 1.388, tube voltage at 120 and 100 kVp, and targeted SD of 9, 14, 20, and 25 respectively. The

AEC function was selected with a range of the tube current between 10 and 400 mA. All other scanning parameters were constantly kept for all data acquisitions as follows: helical scan mode; 80 mm \times 0.5 mm detector configuration; 5 mm slice thickness; rotation time of 0.5 s; scan FOV of 314 mm; 360 mm scan length; Adaptive Iterative Dose Reduction in 3D (ADIR 3D) algorithm was used for the image reconstruction. The real-time displays of volume CT dose index ($CTDI_{vol}$) from the CT monitor were recorded to assess the radiation exposure in each CT protocol.

The $CTDI_{vol}$ was verified for the accuracy, reproducibility, and confidence of using these values. A 32-cm diameter polymethylmethacrylate (PMMA) CT dose index body phantom was used. The measurement was performed by placing a 100 mm pencil-shaped ionization chamber model RaySafe Xi CT detector (Unfors RaySafe Instruments, Billdal, Sweden) at the center and the peripheral positions of the body phantom at the isocenter of the CT bore. The scan parameters were 100 mA, 1 s scan time, 400 mm FOV, and 4×4.0 mm collimation setting for all measurements at kVp 80, 100, 120, and 135 respectively. The PMMA body phantom was scanned three times for each kVp setting. The real-time $CTDI_{vol}$ displayed on CT monitor were recorded and compared with the measured values in percent difference.




Figure 1. **A:** The inner components of the LUNGMAN phantom, **B:** Various sphere sizes of simulated lung nodules of 3, 5, 8, 10, and 12 mm in diameter

Objective image quality

The contrast-to-noise ratio (CNR) was measured by placing the 2 circular regions of interests (ROIs) of similar areas within the nodule and background at the same slice to determine objective image quality (**Figure 2**). Each nodule was measured three times, and the size of ROI was approximately 90% of nodule's boundary. The mean CT number and mean SD within the ROIs were recorded. The SD in the ROI of the background was used to define noise. The CNR is determined using the following formula:

$$CNR = \frac{(CT_{nodule} - CT_{bg})}{SD_{bg}}, \quad (1)$$

where CT_{nodule} and CT_{bg} are the mean CT number of nodule and background respectively, and SD_{bg} is a standard deviation of the background. To eradicate the variance from the nodule location, the CNRs of the simulated nodules of 12, 10, 8, 5, and 3 mm in diameter were grouped.

The CNRs calculated from different scanning parameters were then normalized at targeted SD 9, helical pitch 0.813 and 120 kVp in order to obtain the %CNR in each protocol. The %CNR using the following formula:

$$\%CNR = \frac{(CNR)}{CNR_{(targeted\ SD\ 9,\ pitch\ 0.813,\ 120\ kVp)}} \times 100, \quad (2)$$

where $CNR_{(targeted\ SD\ 9,\ pitch\ 0.813,\ 120\ kVp)}$ is the CNR baseline obtained from the routine-setting CT chest protocol at King Chulalongkorn Memorial Hospital.

Subjective image quality

The evaluation of nodule detection capability was performed by two independent radiologists who have similar experience in chest CT interpretation (TT and PH). They were blinded to the CT scanning parameter techniques, and the CT images were analyzed in randomized order by each reader. Nodule detection capability was graded on a PACS workstation using a five-point rating scale: where 1 denotes *unsatisfactory* (visualize blur of all simulated pulmonary nodules); 2 denotes *poor* (visualize clearly 12 mm, partly visualize 10 mm in diameter); 3 denotes *acceptable* (visualize clearly 10 and 8 mm, partly visualize 5 mm in diameter of simulated nodule); 4 denotes *good* (visualize clearly 5 mm, partly visualize 3 mm in diameter of simulated nodule); and 5 denotes *excellent* (visualize all simulated nodules with a sharp edge).

Statistical analysis

To evaluate interobserver reliability between two radiologists, the kappa (k) test was used for subjective image quality analysis with $k < 0.20$, significant poor agreement; $k = 0.20-0.40$, fair agreement; $k = 0.41-0.60$, moderate agreement; $k = 0.61-0.80$, good agreement; and $k = 0.81-1.0$, perfect agreement [13].

Optimization protocol consideration

The optimal protocol for MDCT chest in this work was selected by considering the targeted SD, beam pitch number, and kVp indicating the lowest possible radiation dose with acceptable image quality for lung nodule detection by two radiologists.



Figure 2. The location of ROIs for measuring the CT number of nodule and background

Results

The results of the radiation dose at various targeted SD, helical pitch, and tube voltage are illustrated in **Table 1**. At the baseline-setting protocol provided the CTDI_{vol} of 5.9 mGy. The highest CTDI_{vol} of 7 mGy was found at targeted SD 9, pitch 0.637 and 120 kVp, whereas the lowest CTDI_{vol} of 0.7 mGy resulted in approximately 10-fold radiation dose reduction was obtained at targeted SD 25, pitch 0.813 and 120 kVp. The variation on targeted SD from 9 to 25 at pitch 0.813 resulted in a radiation dose reduced from 5.9 to 0.7 mGy for 120 kVp, and 5.3 to 0.8 mGy for 100 kVp respectively.

Table 2 demonstrates the percent dose reduction of the CTDI_{vol} on different targeted SD and helical pitch, at 120 and 100 kVp compared with the default setting MDCT chest protocol from manufacturer's scanner at King Chulalongkorn Memorial Hospital. It is found that the increasing of targeted SD from 9 to 25, and helical pitch from 0.637 to 1.388 resulted in the radiation dose decreased by "88% approximately at 120 kVp, and "86% at 100 kVp, meanwhile the optimal protocol at targeted SD 20, pitch 0.813 at 120 kVp can decrease the radiation dose by "81% approximately.

Table 1. Radiation exposure from various targeted SD and helical pitch at 120 and 100 kVp

kVp	Pitch	Targeted SD	CTDI _{vol} (mGy)
120	0.637	9	7
		14	3.1
		20	1.2
		25	0.9
	0.813	9	5.9
		14	2.5
		20	1.1
		25	0.7
	1.388	9	6.5
		14	2.7
		20	1.3
		25	0.9
100	0.637	9	6
		14	2.4
		20	1.3
		25	0.8
	0.813	9	5.3
		14	2.5
		20	1.2
		25	0.8
	1.388	9	6.2
		14	2.4
		20	1.4
		25	0.9

Table 2. The percent reduction of CTDI_{vol} compared with the default setting protocol with increasing targeted SD and helical pitch at 100 and 120 kVp

kVp	Pitch	%CTDI _{vol} reduction			
		Targeted SD			
		9	14	20	25
100	0.637	+1.69	-59.32	-77.97	-86.44
100	0.813	-10.16	-57.63	-79.66	-86.44
100	1.388	+5.08	61.02	77.97	84.75
120	0.637	+18.64	-47.54	-79.66	-84.74
120	0.813	0 (default)	-40.68	-81.35	-88.14
120	1.388	+10.17	-54.24	-77.97	-84.74

A scatter chart representing the average %CNR obtained from all simulated nodule sizes plotted against with the $CTDI_{vol}$ in each of acquisition protocol for 120 and 100 kVp is presented in **Figure 3**. It is demonstrated that the %CNR increases with increasing the radiation dose with $R^2 = 0.867$ for 120 kVp, and $R^2 = 0.852$ for 100 kVp. The %CNR of 120 kVp is higher than 100 kVp for all targeted SD and pitch, because higher kVp increases the number of photons reaching the detectors. The highest %CNR was observed at targeted SD 9, helical pitch 0.637, and 120 kVp, which is provided the highest radiation dose as well. The lowest %CNR was obtained at targeted SD 25, helical pitch 1.388, and 100 kVp.

For the subjective image quality interpreted by radiologists, the 2 readers had substantially interobserver agreement for nodule detection capability (k coefficient = 0.66, $P < 0.05$).

Figure 4 demonstrates an example of cross-sectional thoracic slice of CT chest acquired by the optimal protocol at target SD of 20, pitch 0.813 and 120 kVp. The CT images are displayed by window level/width (WL/WW) setting of 60/360 HU for soft tissue window (**Figure 4A**), and window level/width of -600/1600 HU for lung window (**Figure 4B**) with 5 mm simulated nodule at the 1/3 in peripheral of left lower lobe. Following the consideration for selecting the optimal protocol in our study indicated that this low-dose protocol provided an acceptable image quality of %CNR of 62% with the image scoring of 4 evaluated by 2 radiologists for nodule detection capability 5 mm, and provided the lowest radiation dose compared with other scanning protocols.




Figure 3. The relationship between average %CNR for all lung nodule sizes and radiation doses obtained from 320-slice MDCT with various targeted SD and beam pitch for 120 and 100 kVp




Figure 4. Computed tomography chest images with 5 mm nodule in diameter (arrow) at target SD 20, 10–400 mA, pitch 0.813 and 120 kVp, and displayed with (A) soft tissue window, WL/WW setting of 60/360 Hounsfield Units (HU) and (B) lung window, WL/WW setting of -600/1600 HU

Discussion

Because an academic and social interest in radiation dose reduction for CT examinations without any decrease in diagnostic capability has been growing, the modern MDCT provides a beneficial function such as automatic mA modulation to optimize the radiation dose to patients [14–16]. Targeted SD is a crucial parameter affecting the radiation dose and image quality in MDCT. In the present study, we attempted to improve the protocol for the clinical chest CT study using an appropriate targeted SD and beam pitch to obtain the lowest possible radiation dose, while preserving acceptable image quality.

In the present study, the small focus was set at all acquisitions except at pitch 1.388 on targeted SD 9, so the $CTDI_{vol}$ is higher than the other pitch because the switching from the small focus to the large focus is automatically adjusted the scanning parameters to 120 kVp, 200 mA and 100 kVp, 240 mA. The results of the greater area of the large focus could yield an increased photon flux [15, 16].

For the $CTDI_{vol}$ verification, almost of the values displayed on the monitor were greater than the values measured in PMMA body phantom (Figure 5). The %difference tends to increase with increasing the kVp function. The highest %difference between the displayed and measured $CTDI_{vol}$ values of 8.82% was found at 135 kVp. However, it is reasonably accepted within $\pm 10\%$ using these values [17]. The discrepancy between the measurement and the displayed values result from the uncertainty of chamber position, chamber type and measurement scenario such as

the precision of reading, tube loading, the phantom construction, over scan phenomenon, the detector response in phantoms, and the inaccuracy of laser beam alignment.

A helical pitch 0.637 presented the highest %CNR compared with other pitches for all nodule sizes and targeted SD at both kVp because a low pitch provides low image noise. The %CNR reduced by “50 to “55% when the targeted SD increased from 9 to 25 for all nodule sizes. However, large variation in %CNR of small size nodules, especially nodule sizes 5 and 3 mm diameters were observed. The main variable factors result from the unstable sites of circular ROIs on a very small sized nodule. This also affected to variation of mean CT number of nodule and the SD value of the background.

In a clinical situation, the lung nodules that have a clinical significance are noncalcified nodules (30–40 HU) and ground glass nodules (less than 30 HU) [11, 18]. However, these CT numbers are much lower than CT numbers of simulated nodules designed by manufacturer of the lung man chest phantom in this study (100 HU), which represents the calcified nodules. This is a limitation of this study. Because of the lower CT number of ground glass nodules, noise makes lesion detection more difficult, and ground glass nodules can be missed [18]. Hence, we should be concerned when higher targeted SD and low tube current have been used. As the locations of simulated nodules of all scans were fixed, two readers can recognize the location of simulated nodules. This is another limitation in this study.




Figure 5. Plot of volume of computed tomography detector index ($CTDI_{vol}$) displayed on the monitor (solid line) and measured $CTDI_{vol}$ (dash line) using body technique as a function of kVp for $CTDI_{vol}$ verification of less than 10% discrepancy

Although the appropriate targeted SD was determined by considering the lowest radiation dose with acceptable image quality, optimization protocol should be set up for the benefit of patients in accordance with the clinical purposes. For routine chest CT, the mediastinum, hilar, and pleura should be acquired using a pitch number <1 with high kVp for multi-slice CT. Using targeted SD 20, helical pitch 0.813, and 120 kVp provided the CTDI_{vol} of 1.1 mGy, with an acceptable image quality of %CNR for 62% for nodule detection capability 5 mm with the same agreement interpretation of two radiologists. Although the %CNR reduced by 38% approximately, the simulated nodules can be detected by two readers. This indicated that our low-dose recommended protocol is suitable for using as optimal protocol based on the targeted SD for routine CT chest. However, additional implementation for clinical study in patients is required for further study.

Conclusion

The targeted SD can be revised according to the clinical investigation to keep balancing between patient radiation dose and image quality. Low-dose CT chest protocols in this study can substantially reduce by more than half the relative radiation dose compared with previous routine-setting protocol for 320-slice MDCT while maintaining adequate image quality.

Acknowledgments

The authors thank the radiologists, Drs. Thanisa Tongbai and Patcharee Hongsmatip, for their substantial contributions, analysis and interpretation of the data for this work. This work was presented as an oral presentation at the 13th Asia-Oceania Congress of Medical Physics (AOCMP) and 11th South-East Asian Congress of Medical Physics (SEACOMP) December 12–14, 2013 in Singapore, and published in the conference proceedings of AOCMP & SEACOMP 2013 by Nuntue C, Krisanachinda A, Khamwan K, Tongbai T, Hongsmatip P. Optimization of 320 MDCT using beam pitch and tube current modulation for proper Target SD on lung nodule –Chest protocol : Phantom study. Sincere thanks Professor Franco Milano and Associate Professor Sivalee Suriyapee for their invaluable advice. We also thank Mrs. Phetchareeya Suwanpradit, Mrs. Walaiporn Suksancharoen, and other technologists at the Unit of Computed Tomography, King Chulalongkorn Memorial Hospital, Bangkok, Thailand for their kind support in this work.

Conflict of interest statement

The authors declare that there is no conflict of interest in this research.

References

1. Kubo T, Ohno Y, Kauczor HU, Hatabu H. Radiation dose reduction in chest CT—review of available options. *Eur J Radiol.* 2014; 83:1953-61.
2. Kubo T, Lin PJ, Stiller W, Takahashi M, Kauczor HU, Ohno Y, et al. Radiation dose reduction in chest CT: a review. *AJR.* 2008; 190:335-43.
3. Arapakis I, Efsthathopoulos E, Tsitsia V, Kordolaimi S, Economopoulos N, Argentos S, et al. Using “iDose4” iterative reconstruction algorithm in adults’ chest–abdomen–pelvis CT examinations: effect on image quality in relation to patient radiation exposure. *Br J Radiol.* 2014; 87:20130613.
4. Bankier AA, Tack D. Dose reduction strategies for thoracic multidetector computed tomography: background, current issues, and recommendations. *J Thorac Imaging.* 2010; 25:278-88.
5. Trinavarat P, Kritsaneepaiboon S, Rongviriyapanich C, Visrutaratna P, Srinakaran J. Radiation dose from CT scanning: can it be reduced? *Asian Biomed.* 2011; 5:13-21.
6. Ohno Y, Takenaka D, Kanda T, Yoshikawa T, Matsumoto S, Sugihara N, et al. Adaptive iterative dose reduction using 3D processing for reduced- and low-dose pulmonary CT: comparison with standard-dose CT for image noise reduction and radiological findings. *AJR.* 2012; 199:477-85.
7. McCollough CH, Bruesewitz MR, Kofler JM. CT dose reduction and dose management tools: overview of available options. *Radiographics.* 2006; 26:503-12.
8. Li Q, Yu H, Zhang L, Fan L, Liu SY. Combining low tube voltage and iterative reconstruction for contrast-enhanced CT imaging of the chest—initial clinical experience. *Clin Radiol.* 2013; 68:e249-53.
9. Lee CH, Goo JM, Ye HJ, Ye SJ, Park CM, Chun EJ, et al. Radiation dose modulation techniques in the multidetector CT era: from basics to practice. *Radiographics.* 2008; 28:1451-59.
10. Kalra MK, Maher MM, Toth TL, Schmidt B, Westerman BL, Morgan HT, et al. Techniques and applications of automatic tube current modulation for CT. *Radiology.* 2004; 233:649-57.
11. Matsumoto K, Ohno Y, Koyama H, Kono A, Inokawa H, Onishi Y, et al. 3D automatic exposure control for 64-detector row CT: radiation dose reduction in chest phantom study. *Eur J Radiol.* 2011; 77:522-27.

12. Namasivayam S, Kalra MK, Pottala KM, Waldrop SM, Hudgins PA. Optimization of Z-axis automatic exposure control for multidetector row CT evaluation of neck and comparison with fixed tube current technique for image quality and radiation dose. *Am J Neuroradiol*. 2006; 27:2221-25.
13. Svanholm H, Starklint H, Gundersen HJ, Fabricius J, Barlebo H, Olsen S. [Reproducibility of histomorphologic diagnoses with special reference to the kappa statistic](#). *APMIS*. 1989; 97:689-98.
14. Goldman LW. Principles of CT: radiation dose and image quality. *J Nucl Med Technol*. 2007; 35:213-25.
15. Hamberg LM, Rhea JT, Hunter GJ, Thrall JH. Multidetector row CT: radiation dose characteristics. *Radiology*. 2003; 226:762-72.
16. McNitt-Gray MF. AAPM/RSNA physics tutorial for residents: topics in CT. Radiation dose in CT. *Radiographics*. 2002; 22:1541-53.
17. International Atomic Energy Agency. Quality Assurance program for computed tomography: diagnostic and therapy application. IAEA Human Health Series No 19. Vienna: International Atomic Energy Agency; 2012.
18. Funama Y, Awai K, Liu D, Oda S, Yanaga Y, Nakaura T, et al. Detection of nodules showing ground-glass opacity in the lungs at low-dose multidetector computed tomography: phantom and clinical study. *J Comput Assist Tomogr*. 2009; 33:49-53.